

ASB Standard 098, First Edition  
2020

**Standard for Mass Spectral Data Acceptance in Forensic  
Toxicology**

DRAFT



## Standard for Mass Spectral Data Acceptance in Forensic Toxicology

ASB Approved Xxxxxx 2020

ANSI Approved Xxxxxxx 2020



Academy Standards Board  
410 North 21<sup>st</sup> Street  
Colorado Springs, CO 80904

This document may be downloaded for free at: [www.asbstandardsboard.org](http://www.asbstandardsboard.org)

*This document is provided by the AAFS Standards Board for free. You are permitted to print and download the document and extracts from the document for your own use, provided that:*

- *you do not modify this document or its related graphics in any way;*
- *you do not use any illustrations or any graphics separately from any accompanying text; and,*
- *you include an acknowledgement alongside the copied material noting the AAFS Standards Board as the copyright holder and publisher.*

*You expressly agree not to reproduce, duplicate, copy, sell, resell, or exploit for any commercial purposes, this document or any portion of it. You may create a hyperlink to [www.asbstandardsboard.org](http://www.asbstandardsboard.org) to allow persons to download their individual, free copy of this document. Your hyperlink must not portray AAFS, the AAFS Standards Board, this document, our agents, associates and affiliates in an offensive manner, or be misleading or false. You may not use our trademarks as part of your link without our written agreement for you to do so.*

*The AAFS Standards Board retains the sole right to submit this document to any other forum for any purpose.*

*Certain commercial entities, equipment or materials may be identified in this document to describe a procedure or concept adequately. Such identification is not intended to imply recommendations or endorsement by the AAFS or the AAFS Standards Board, nor is it intended to imply that the entities, materials, or equipment are necessarily the best available for the purpose.*

*This document is copyrighted © by the AAFS Standards Board, LLC. 2020 All rights are reserved.  
410 North 21st Street, Colorado Springs, CO 80904, [www.asbstandardsboard.org](http://www.asbstandardsboard.org).*

## Foreword

During the last several decades, mass spectrometry has replaced traditional, less specific techniques such as flame ionization, nitrogen-phosphorus, electron-capture, ultraviolet and fluorescence detection as the preferred technology for the confirmation of drugs, drug metabolites, relevant xenobiotics, and endogenous analytes in forensic toxicology. Although criteria for the acceptance of mass spectrometry data have been promulgated in regulated areas of forensic toxicology, none have been universally applied by practicing forensic toxicologists.

This document addresses this gap by providing minimum standards of practice for the acceptance of mass spectral data used in all forensic toxicology laboratories. Specifically, this standard focuses on minimum criteria for mass spectral data acquired using a nominal or high resolution mass spectrometer that utilizes ionization processes such as electron ionization, chemical ionization, electrospray ionization, or atmospheric pressure chemical ionization.

This standard was prepared and finalized as a standard by the Toxicology Consensus Body of the AAFS Standards Board. This document was originally conceived by the Scientific Working Group for Toxicology (SWGTOX) and further developed by the Toxicology Subcommittee of the Organizational Scientific Area Committee (OSAC) for Forensic Science.

All hyperlinks and web addresses shown in the document are current as of the publication date of this standard.

**Keywords:** *Mass spectrometry, mass spectral data.*

**Table of Contents**

1 Scope.....

2 Normative References .....

3 Terms and Definitions .....

4 Requirements .....

Annex A (informative) Bibliography .....

DRAFT

# Standard for Mass Spectral Data Acceptance in Forensic Toxicology

## 1 Scope

This document provides criteria for the acceptance of mass spectral analyses of small molecules (compounds with an atomic weight of less than 800 daltons) in laboratories conducting any of the following forensic toxicology subdisciplines: postmortem forensic toxicology, human performance toxicology (e.g., drug-facilitated crimes and driving-under-the-influence of alcohol or drugs), non-regulated employment drug testing, court-ordered toxicology (e.g., probation and parole, drug courts, child services), and general forensic toxicology (non-lethal poisonings or intoxications).

The document provides minimum requirements for acquiring data on single- or multiple-stage mass spectrometers using nominal or high-resolution mass spectrometers. It also provides instruction on the evaluation of mass spectral data when conducting acquisitions in full-scan mode, selected ion monitoring, multiple-stage analyses, or when using high-resolution mass analyzers.

Criteria, requirements and instructions in this document are not intended for the area of breath alcohol toxicology. Further, it is not intended to address the use of matrix assisted laser desorption, inductively coupled plasma, or ion mobility mass spectrometry. It is also not intended to provide criteria for analyte identification in forensic toxicology laboratories.

## 2 Normative References

The following references are indispensable document for the application of the standard. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ANSI/ASB Standard 036, *Standard Practices for Method Validation in Forensic Toxicology*, First Edition

ASB Standard 054: *Standard for Quality Control Program in Forensic Toxicology*, First Edition  
*[document not yet published, will not be included if it is not published]*

ASB Standard 113, *Standard for Identification Criteria in Forensic Toxicology*. First Edition  
*[document not yet published, will not be included if it is not published]*

## 3 Terms and Definitions

For purposes of this document, the following definitions apply.

### 3.1

#### **base peak**

Most abundant ion in the mass spectrum. In plotting the mass spectrum, all other ions are normalized to the base peak<sup>[13]</sup>.

### 3.2

#### **diagnostic ion**

An ion whose formation reveals structural or compositional information about the target analyte<sup>[14]</sup>.

### **3.3 high resolution HR**

In this document, it refers to a MS instrument that can give at least 10,000 nominal mass resolving power at full width of the peak at half its maximum height (FWHM) for the compound of interest.<sup>[1]</sup>

### **3.4 ionization**

The physicochemical process of producing a gas phase ion. In the mass spectrometer this typically occurs within the ion source. Several mechanisms of ionization exist such as chemical ionization and electron ionization.

### **3.5 mass spectrometry MS**

Study of matter through the formation of gas-phase ions that are characterized using mass spectrometers by their mass, charge, structure, and/or physio chemical properties<sup>[14]</sup>.

### **3.6 monoisotopic mass**

Exact mass of an ion or molecule calculated using the mass of the most abundant isotope of each element<sup>[14]</sup>.

### **3.7 MS<sup>n</sup>**

Multiple-stage mass spectrometry experiments designed to record product ion spectra where n is the number of product ion stages (nth-generation product ions)<sup>[14]</sup>.

### **3.8 multiple reaction monitoring MRM**

Application of selected reaction monitoring to multiple product ions from one or more precursor ions<sup>[14]</sup>.

### **3.9 nominal mass analyzer**

In this document, it refers to a MS instrument that can achieve resolution of <5000.

### **3.10 precursor ion**

Ion that reacts to form product ions or undergoes specified neutral losses<sup>[14]</sup>.

### **3.11 product ion**

Ion formed as the product of a reaction involving a precursor ion<sup>[14]</sup>.

### **3.12 reference material**

Material, sufficiently homogenous and stable with respect to one or more specified properties, which has been established to be fit for its intended use in a measurement<sup>[13]</sup>.

### 3.13 relative abundance

The abundance of an ion produced in relation to the abundance of the base peak<sup>[21]</sup>.

### 3.14 resolution

In a mass spectrum, the observed  $m/z$  value divided by the smallest difference  $\Delta(m/z)$  for two ions that can be separated:  $(m/z)/\Delta(m/z)$ <sup>[14]</sup>.

### 3.15 selected ion monitoring SIM

Operation of a mass spectrometer in which the abundances of ions of one or more specific  $m/z$  values are recorded rather than the entire mass spectrum<sup>[14]</sup>.

### 3.16 selected reaction monitoring SRM

Data acquired from one or more specific product ions corresponding to  $m/z$  selected precursor ions recorded via two or more stages of mass spectrometry<sup>[14]</sup>.

### 3.17 tandem mass spectrometry MS/MS

Acquisition and study of the spectra of the product ions or precursor ions of  $m/z$  selected ions, or of precursor ions of a selected neutral mass loss<sup>[14]</sup>.

## 4 MS Analysis Criteria Requirements

### 4.1 General Rules

MS analysis may be performed by several approaches including single-stage or multiple-stage instruments at nominal and high resolution. Data may be acquired using full-scan or by monitoring selected ions and are used to establish evidence of analyte identification, per ASB Standard 113: *Standard for Identification Criteria in Forensic Toxicology* [document not yet published].

Selected ions shall be diagnostic ions. Ions that represent a loss of derivatizing agent or loss of water shall not be considered as structurally relevant. Intact isotopomers do not allow for structural elucidation by MS and shall not be used as diagnostic ions. Adducts (such as dimers) shall not be considered as diagnostic ions because no meaningful information is obtained from their fragmentation.

Instrument operating conditions (e.g. tune parameters, scan range, monitored ions, ionization energies, dwell time) for MS analysis shall be the same as used during method validation studies and follow the requirements of ANSI/ASB Standard 036: *Practices for Method Validation in Forensic Toxicology*.

Method performance shall be continuously monitored with a rigorous quality assurance and quality control (QA/QC) program according to ASB Standard 054: *Standard for Quality Control Program in Forensic Toxicology* [document not yet published]. Further, documented regular instrument

maintenance shall follow recommendations of the manufacturer and described in the appropriate laboratory procedures.

## 4.2 Single-stage Mass Analysis Using a Nominal Mass Analyzer

### 4.2.1 Full-scan Acquisition for Screening

The following criteria shall be met when using single-stage nominal mass analyzer for screening purposes in full-scan mode.

- a) The scan range shall be the same as that established during method validation and include all diagnostic ions. The scan range should be set to at least 50 m/z above the highest anticipated fragment.
- b) A minimum of a single diagnostic ion shall be monitored.

### 4.2.2 Full-scan Acquisition for Confirmation

The following criteria shall be met when using single-stage nominal mass analyzer for confirmation purposes in full-scan mode.

- a) The scan range shall be the same as that established during method validation and include all diagnostic ions. The scan range should be set to at least 50 m/z above the highest anticipated fragment.
- b) No ions shall be present at a relative abundance equal to or greater than 50% that are not present in the reference material spectrum.
- c) When background subtractions are performed, they shall be documented.
- d) Ratios of diagnostic ions:
  - 1) Shall agree with those calculated from a reference material given the tolerances shown in Table 1; or
  - 2) The spectrum shall be compared to a library by a forward and reverse spectral match algorithm and be above a documented, pre-defined threshold (match factor).

**Table 1—Maximum Acceptable Ion Ratio Limits**

(adapted from references 4,7, and 21)

Relative Abundance	Electron Impact Ionization	All Other Ionization Techniques
Greater than 50%	20%	20%
20 to 50%	20%	25%
Less than 20%	20%	30%

### 4.2.3 SIM for Screening

When using a single-stage nominal mass analyzer for screening purposes in SIM mode, a minimum of a single diagnostic ion shall be monitored.

### 4.2.4 SIM for Confirmation

When using a single-stage nominal mass analyzer for confirmation purposes in SIM mode, diagnostic ion ratios shall agree with those calculated from a concurrently analyzed reference material given the tolerance shown in Table 1.

## 4.3 Multiple-stage Mass Analysis Using Nominal Mass Analyzer (MS/MS or MS<sup>n</sup>)

### 4.3.1 Product Ion Scan Acquisition for Screening

The following criteria shall be met when using a multiple-stage nominal mass analyzer for screening purposes in product ion scan mode.

- a) The precursor mass shall have unit resolution or better.
- b) The scan range shall be the same as that established during method validation. The upper end of the scan range for product ion spectra should be set above the precursor ion  $m/z$ .
- c) A minimum of a single transition of a diagnostic ion shall be monitored.

### 4.3.2 Product Ion Scan Acquisition for Confirmation

The following criteria shall be met when using a multiple-stage nominal mass analyzer for confirmation purposes in product ion scan mode.

- a) The precursor mass shall have unit resolution or better.
- b) The scan range for product ion spectra shall be the same as that established during method validation. The upper end of the scan range for product ion spectra should be set above the precursor ion  $m/z$ .
- c) A minimum of a single transition of a diagnostic ion shall be monitored.
- d) No ions shall be present at a relative abundance equal to or greater than 50% that are not present in the reference material spectrum.
- e) Ratios of diagnostic ions:
  - 1) shall agree with those calculated from a reference material given the tolerances shown in Table 1; or
  - 2) the spectrum shall be compared to a library by a forward and reverse spectral match algorithm and be above a pre-defined threshold.

### **4.3.3 MRM Acquisition for Screening**

The following criteria shall be met when using a multiple-stage nominal mass analyzer for screening purposes in MRM mode.

- a) The precursor mass shall have unit resolution or better.
- b) A minimum of a single transition of a diagnostic ion shall be monitored.

## **4.4 Single-stage Mass Analysis Using High-resolution Analyzer**

### **4.4.1 Single-stage Mass Analysis with High-resolution Acquisition for Screening**

When using a single-stage high-resolution mass analyzer for screening purposes, a minimum of a single diagnostic ion shall be monitored.

### **4.4.2 Single-stage Mass Analysis with High-resolution Acquisition for Confirmation**

The following criteria shall be met when using a single-stage high-resolution mass analyzer for confirmation purposes.

- a) The scan range shall be the same as that established during method validation and include all diagnostic ions. The scan range should be set to at least 50 m/z above the highest anticipated fragment.
- b) No ions shall be present at a relative abundance equal to or greater than 50% that are not present in the reference material spectrum.
- c) The spectrum shall contain the molecular species within 5 ppm of the theoretical monoisotopic mass.
- d) Ratios of diagnostic ions:
  - 1) shall agree with those calculated from a reference material given the tolerances shown in Table 1; or
  - 2) the spectrum shall be compared to a library by a forward and reverse spectral match algorithm and be above a pre-defined threshold.

## **4.5 Multiple-stage Mass Analysis Using A High-resolution Analyzer (MS/MS or MS<sub>n</sub>)**

### **4.5.1 Multiple-stage Mass Analysis with High-resolution Acquisition for Screening**

When using a multiple-stage high-resolution mass analyzer for screening purposes, a minimum of a single diagnostic ion shall be monitored.

### **4.5.2 Multiple-stage Mass Analysis with High-resolution Acquisition for Confirmation**

The following criteria shall be met when using a multiple-stage high-resolution mass analyzer for confirmation purposes.

- a) The precursor mass shall have unit resolution or better.
- b) The scan range shall be the same as that established during method validation and include all diagnostic ions. The scan range should be set to at least 50 m/z above the highest anticipated fragment.
- c) No ions shall be present at a relative abundance equal to or greater than 50% that are not present in the reference material spectrum.
- d) The spectrum must contain the molecular species within 5 ppm of the theoretical monoisotopic mass.
- e) Ratios of diagnostic ions:
  - 1) shall agree with those calculated from a reference material given the tolerances shown in Table 1; or
  - 2) the spectrum shall be compared to a library by a forward and reverse spectral match algorithm and be above a pre-defined threshold.

DRAFT

## Annex A (informative)

### Bibliography

The following information provides a list of the literature resources that may assist the DNA technical leader in defining the breadth and scope of the materials to be reviewed by the trainee. This list is not meant to be all inclusive. The laboratory shall develop a list tailored to its specific needs. Updated references shall be added to the laboratory's list as new methods or technologies are incorporated into the laboratory's protocols.

- 1] Acceptance Criteria for Confirmation of Identity of Chemical Residues using Exact Mass Data for the FDA Foods and Veterinary Medicine Program US Food & Drug Administration Office of Foods and Veterinary Medicine September 2015
- 2] Berendsen, B.J., L.A. Stolker, and M.W. Nielen, . "The (Un) Certainty of Selectivity in Liquid Chromatography Tandem Mass Spectrometry." *Journal of the American Society for Mass Spectrometry*, 2012. 24, 154-163.
- 3] Clinical and Laboratory Standards Institute, & International Federation of Clinical Chemistry and Laboratory Medicine (2007). *Mass Spectrometry in the Clinical Laboratory: General Principles and Guidance; Approved Guideline (CLSI document C50-A)*. Wayne, Pennsylvania: Clinical and Laboratory Standards Institute.
- 4] Commission of the European Communities (2002). Commission Decision of 12 August 2002 implementing Council Directive 96/23/EC concerning the performance of analytical methods and the interpretation of results (notified under document number C3044) (Text with EEA relevance) (657). Official Journal of the European Communities.
- 5] Cooper, G. A., Paterson, S., & Osselton, M. D. (2010). The United Kingdom and Ireland Association of Forensic Toxicologists Forensic toxicology laboratory guidelines. *Science and Justice*, 50, 166-176.
- 6] Couchman, L., & Morgan, P. (2011). LC-MS in analytical toxicology: some practical considerations. *Biomedical Chromatography*. 25, 100-123.
- 7] Guidance Document for Laboratories and Inspectors, National Laboratory Certification Program, Research Triangle Park, NC, 2002.
- 8] Habib Jiwan, J. L., Wallemacq, P., & Herent, M. F. (2010). HPLC-high resolution mass spectrometry in clinical laboratory? *Clinical Biochemistry*, 44, 136-147.
- 9] Hoyt, D. W., Finnigan, R. E., Nee, T., Shults, T. D., & Butler, T. J. (1987). Drug Testing in the Workplace-Are Methods Legally Defensible? A Survey of Experts, Arbitrators, and Testing Laboratories. *Journal of the American Medical Association*, 258(4), 504-509.
- 10] Lehotay, S. J., Mastovska, K., Amirav, A., Fialkov, A. B., Alon, T., Martos, P. A., Kok, A. D., & Fernandez-Alba, A. R. (2008). Identification and confirmation of chemical residues in food by

- chromatography-mass spectrometry and other techniques. *Trends in Analytical Chemistry*, 27(11), 1070-1090.
- 11] Maurer, H. H. (2005). Advances in analytical toxicology: the current role of liquid chromatography-mass spectrometry in drug quantification in blood and oral fluid. *Analytical and Bioanalytical Chemistry* 381, 110-118.
  - 12] Peters, F. T., & Remane, D. (2012). Aspects of matrix effects in applications of liquid chromatography-mass spectrometry to forensic and clinical toxicology—a review. *Analytical and Bioanalytical Chemistry* 403, 2155-2172.
  - 13] OSAC Preferred Terms. Retrieved from <http://lexicon.forensicosac.org/Term/Home/Index>. Accessed March 14, 2018
  - 14] *Pure Appl. Chem.*, Vol. 85, No. 7, pp. 1515–1609, 2013. <http://dx.doi.org/10.1351/PAC-REC-06-04-06>
  - 15] RTI International (2010). The National Laboratory Certification Program Manual for Urine Laboratories, October (rev. July 2011).
  - 16] Sauvage, F. L., Gulier, J. M., Lachatre, G., & Marquet, P. (2008). Pitfalls and Prevention Strategies for Liquid Chromatography-Tandem Mass Spectrometry in the Selected Reaction Monitoring Mode for Drug Analysis. *Clinical Chemistry*. 54, 1519-1527.
  - 17] Scientific Working Group for the Analysis of Seized Drugs (SWGDRUG) (2011). Scientific Working Group for the Analysis of Seized Drugs (SWGDRUG) Recommendations. Retrieved from <http://www.swgdrug.org/>
  - 18] Scientific Working Group for Forensic Toxicology (SWGTOX) Standard Practices for Method Validation in Forensic Toxicology, *Journal of Analytical Toxicology*, Volume 37, Issue 7, 1 September 2013, Pages 452–474, <https://doi.org/10.1093/jat/bkt054>
  - 19] U.S. Department of Health and Human Services Food and Drug Administration, Center for Drug Evaluation and Research, & Center for Veterinary Medicine (2001). Guidance for Industry Bioanalytical Method Validation. Retrieved from <http://www.fda.gov/cder/guidance/index.htm>
  - 20] U.S. Department of Health and Human Services Food and Drug Administration, Center for Veterinary Medicine (2001). Guidance for Industry Mass Spectrometry for Confirmation of the Identity of Animal Drug Residues US FDA/CVM. May 1, 2003 Retrieved from <https://www.fda.gov/downloads/AnimalVeterinary/GuidanceComplianceEnforcement/GuidanceforIndustry/ucm052658.pdf>
  - 21] WADA (2003). Identification Criteria for Qualitative Assays Incorporating Chromatography and Mass Spectrometry (TD2003IDCR).

DRAFT



Academy Standards Board  
410 North 21st Street  
Colorado Springs, CO 80904

[www.asbstandardsboard.org](http://www.asbstandardsboard.org)